Lab Chip

ARTICLE

Received 00th January 20xx,

Pneumatically actuated cell-stretching array platform to engineer cell pattern *in vitro*

Harshad Kamble,^a Raja Vadivelu,^b Mathew Barton,^{c,d} Muhammad J. A. Shiddiky,^b Nam-Trung Nguyen^{a*}

Accepted 00th January 20xx

DOI: 10.1039/x0xx00000x

www.rsc.org/

A cellular response to mechanical stimuli is well-known phenomena known as a mechanotransduction. It is widely accepted that mechanotransduction plays an important role in cell alignment which is critical for cell homeostasis. Although many approaches have been developed in recent years to study the effect of external mechanical stimuli on cell behaviour, most of them have not explored the ability of the mechanical stimuli to engineer cell alignment to obtain patterned cell culture. This paper introduces a simple, yet effective pneumatically actuated 4x2 cell stretching array for concurrently inducing a range of cyclic normal strain onto cell cultures to achieve predefined cell alignment. We utilised ring-shaped normal strain pattern to demonstrate the growth of *in vitro* patterned cell culture with predefined circumferential cellular alignment. Furthermore, to ensure the compatibility of the developed cell stretching platform with general tools and existing protocols, dimensions of the developed cell-stretching platform follow the standard F bottom 96 well plate. In this study we report principle design, simulation and characterisation of the cell-stretching platform with the preliminary observations using fibroblast cells. Our experimental results of cytoskeleton reorganisation such as perpendicular cellular alignment of the cells to the direction of normal strain are consistent with those reported in the literature. After two hours of stretching, the circumferential alignment of fibroblast cells confirms the capability of the developed system to achieve a patterned cell culture. The cell-stretching platform reported is potentially a useful tool for drug screening in 2D mechanobiology experiments, tissue engineering and regenerative medicine.

1 Introduction

Cells within the human body are responsive to the surrounding mechanical stimuli.¹⁻⁴ The ability of cells to sense and respond to mechanical stimuli is known as a mechanotransduction.⁵⁻⁷ At the cellular level, mechanical stimulus acts as a key driver in regulating cellular functions. It is well known that such mechanical stimuli induce biochemical responses, which are important to maintain normal cellular activities such as cytoskeleton organisation, cellular orientation and pattern formation during tissue maturation and regeneration.⁸⁻¹² Thus, the complex mechanotransduction signalling process is critical for cells to maintain homeostasis.

For instance, many previous studies reported that abnormal mechanotransduction function within the cells such as maladaptive remodelling of cytoskeleton may disturb the

cellular activities and lead to diseases.^{13,14} Furthermore, some studies also claimed that external forces acting on the cells also can influence the cell functionality by regulating the mechanical properties of ECM.¹⁵⁻¹⁷

Thus, better understanding of the mechanotransduction signalling process is critical as it could provide an insight on the role of mechanical stimuli in guiding cytoskeleton reorganization and cell reorientation.

Indeed, various methods have been developed to apply external mechanical stimuli *in-vitro* to observe cellular response. Common techniques to induce mechanical force onto the cell culture use existing tools such as atomic force microscope (AFM), micropipette or tweezers.¹⁸⁻²¹ In recent years, more sophisticated cell-stretching designs have been introduced into the commercial market. Flexcell (Flexcell International Corporation), ElectroForce (Bose Corporation) and Strex Systems for cell Stretching (STREX Inc.) are some of the widely accepted commercial platforms. Many cell-stretching studies included these commercial platforms due to its well characterised parameters and ease of operation.²²⁻²⁴

Apart from these commercial approaches, many custom made cell stretching platforms also have been reported. Many of them incorporates stretching of the elastic membrane with

^a QLD Micro- and Nanotechnology Centre, Nathan Campus, Griffith University, 170 Kessels Road, Brisbane, QLD 4111, Australia.

^b School of Natural Sciences, Nathan Campus, Griffith University, 170 Kessels Road, Brisbane, QLD 4111, Australia.

^c HealthTech-X, Menzies Health Institute Queensland, Griffith University, Gold Coast, QLD, Australia.

^d Clem Jones Centre for Neurobiology and Stem Cell Research, Don Young Road Nathan Qld 4111

cell culture by the means of electromagnetic, pneumatic, optical or other micro actuation techniques.²⁵⁻³⁰ For instance, Huang and Nguyen (2013) developed a multilayered PDMS device with pneumatic actuators to deform the elastic membrane and introduce normal strain onto the cells cultured on the membrane. Kamotani, *et al.* (2008) reported an array of PDMS microcells with deformable membrane, placed over piezoelectrically actuated pins. Furthermore, Chang, *et al.* (2013)³³ utilised a motor actuated translation stage to stretch the deformable substrate in one direction and to introduce normal strain onto the cells cultured on the substrate. Electro-thermal, electrostatic and di-electrophoresis actuations have also been explored for cell stretching.³⁴⁻³⁶ Although, all these existing cell stretching approaches have particular advantage for their specific applications, most of these studies were focused on investigating the effect of the normal strain on cellular alignment. For instance, Wang, *et al.* (2015) used hydrogel of a linear elastic property to develop programmable normal strain gradients. Interestingly, authors observed that cell realigns perpendicular to the substrate normal strain, and cell realignment was approximately linear to the applied strain. In another study Chagnon-Lessard, *et al.* (2017) showed the capability of the cells to avoid normal strain gradient under highly non-uniform strain field and confirmed the cells capability to respond not only to the strain amplitude but also to the strain direction. Considering such substantial results it was interesting to note that very few studies have actually explored the possibility of active cell manipulation using external mechanical stimuli for achieving patterned cell culture with predefined cellular alignment. Thus, a simple and effective method compatible with existing pre-clinical tools and protocols is urgently needed for engineering patterned cell culture which could be useful in the clinical drug screening for 2D mechanobiological experiments.

This paper presents a pneumatically actuated cell stretching array platform (Fig. 1(a)), which can apply a range of normal strains onto the cell cultures to achieve predefined *in vitro* circumferential cellular alignment. The dimensions of the cell-stretching platform were maintained same as the standard 96 well plates. Thus, the developed cell-stretching platform is compatible with the general pre-clinical tools and protocols. The compatibility could be critical for designing high-throughput cell-stretching assays.

To prove the concept, the presently developed cell-stretching platform incorporates an array of 4x2 wells. The wells in the each column was utilised for introducing different predefined ring shaped normal strain gradient onto the cultured cells. As a first demonstration, we describe the design, simulation, fabrication and characterisation of the developed cell stretching platform in detail. Further, for confirming the suitability of the developed platform and to obtain patterned cell culture with predefined circumferential alignment, we considered fibroblast cells for this study.

Fibroblast is a unique cell type which shares the plastic behaviour of mesenchymal phenotype and can be found in abundance within a human body. Hence, it is considered as a prime candidate to understand the role of mechanotransduction in altering the cell behaviour. Many studies have shown the capability of the fibroblast cells to recognise and respond to the external cyclic normal strain and reorient perpendicular to the normal strain direction.³⁸⁻⁴²

Thus, respecting existing literature we hypothesised that the cyclic stretching of the fibroblast cells with a ring shaped output stretching pattern may facilitate patterned fibroblast cell culture with circumferential cellular alignment.

2 Materials and methods

2.1 Device design and working principle

The pneumatically actuated cell-stretching array platform was designed and characterised to achieve patterned cell culture with circumferential alignment. Furthermore, the platform was particularly designed to induce different pre-defined normal strain onto the cells cultured in each well to establish the relationship between normal strain amplitude and cellular alignment, Figure 1(b,c). The developed platform provides advantages such as ease of fabrication, ease of imaging, compatibility with generic pre-clinical tools and predefined normal strain pattern. The array design makes it suitable for the high throughput cellular response assays, Figure 1(d).

Figure 1(e) depicts the actual image of the cell-stretching platform. The platform includes three PDMS layers: (i) the lower layer consisting of the microfluidic channels and the ring shaped pneumatic chamber, (ii) the middle layer consisting of a 200- μ m deformable PDMS membrane bonded with lower layer and finally, (iii) the upper layer consisting of the cell culture wells which are aligned with the output patterns to facilitate cell seeding. For the current study, output stretching pattern was varied for wells in each column to generate a range of normal strain magnitudes with a single device. Each normal strain pattern consists of a toroidal structure with a fixed outer diameter of 6.39 ± 0.01 mm and variable inner radius (R_i). The variable nature of the inner radius allows achieving variable normal strain amplitude onto the deformable membrane upon actuation.

Figure 1(d) shows the deformable ring-shaped output pattern 1 (OP1) with the inner radius $R_{i1}=2$ mm, deformable output pattern 2 (OP2) with the inner radius $R_{i2}=1.5$ mm and the deformable output pattern 3 (OP3) with the inner radius $R_{i3}=1$ mm. The dimension of the output patterns was maintained according to the F-bottom standard 96 well polystyrene microplates dimension (Greiner Bio-One International GmbH) to maintain the compatibility with generic tools. Similar to a standard 96 well plate, the centre to centre distance between each pattern was maintained at 9.02 ± 0.01 mm with a fixed outer diameter of 6.39 mm for all

output patterns (OPs). Further, the depth of the cell culture wells was maintained at 5 ± 0.5 mm.

We used a pressure controller (Elveflow®) to actuate the OPs by providing positive pressure of a given magnitude and frequency to the inlets of the cell stretching platform. Upon actuation, the membranes over the output patterns inflated and deformed into semi toroid shapes which subsequently induced normal strain onto the cells cultured on the deformable OPs.

2.2 Design and optimisation

To assess the feasibility of the developed cell stretching platform and predict the system behaviour, a finite element analysis (FEA) model of the platform was formulated in COMSOL Multiphysics 5.2. In the optimisation phase, we optimised the geometry of the platform by varying one geometric parameter and maintaining the other parameters constant. The variable inner radius for the output pattern was one of the key parameter for the optimisation, as the possible normal strain amplitudes can be determined. Assuming a constant pressure, the reference geometry was adapted in COMSOL with the total volume of $9 \times 9 \times 10.2$ mm³.

The outer diameter was fixed at 6.39 mm, and the inner diameter and height were maintained at 3 mm and 0.1 mm respectively. Furthermore, the height of the channel and the membrane thickness were kept at 0.1 mm and 0.2 mm respectively. The material was selected as polydimethylsiloxane (PDMS) for the entire geometry with the Young's modulus as 750 Kpa and Poisson's ratio as 0.49 for PDMS mixed with crosslinker in 10 to 1 volume ratio. In the next step, necessary boundary conditions were implemented for the reference FEA model by allowing free deformation of the output pattern and fixing all other parts. A positive pressure of 80 mbar was applied to the membrane to represent the pneumatic actuation. Finally, utilising parametric sweep function in COMSOL, the inner radius of the output pattern was varied from 0.25 mm to 5.5 mm with a step of 0.25 mm to obtain the corresponding maximum planar displacement of the membrane.

Figure 2 shows the membrane displacement as function of the inner radius. The simulation provided a wide range of normal strain amplitudes that could be achieved with the current cell stretching platform. Such a wide range allows for establishing the relationship between cellular alignment and applied normal strain. In the current study, we kept the inner diameter for the OP1, OP2 and OP3 at 3 mm, 2 mm and 1 mm respectively. The insets in Figure 2 show the simulation results for the selected OPs.

2.3 Fabrication process

The software CleWin 3.0 was utilised to design the cell-stretching platform. The optimised design was plotted on the 28 mmx22 mm plastic mask for pattern replication. The fabrication of the multilayered stretching platform mainly

involved the soft lithography technique. Figure 3 illustrates the basic fabrication steps. The process comprised of three major steps.

In the first step, a clean and pre-baked (80 °C for 30 min) silicon wafer was spin coated with a 100-μm layer of photoresist (SU-8 50) at 3,300 rpm for 30 sec. A series of pre-baking steps (65 °C for 10 min, 95 °C for 30 min, 50 °C for 30 mins) were performed to avoid SU-8 cracking. Subsequent UV exposure (9.8 mW/cm² for 35 to 40 seconds) allowed transferring the device pattern to the wafer. Finally, the SU-8 was post baked (80 °C for 30 mins) and the mask was developed in 1-methoxy-2-propanol acetate for 15 minutes. The same procedure was followed to prepare the master mould for the upper PDMS layer with the cell culture wells.

In step two, a clear degassed PDMS and cross linker (Sylgard 184 elastomer kit, Dow Corning) mixture (10:1 volume ratio) was poured onto the master moulds, so that a thickness of 2-2.5 mm and 5 mm was achieved for the lower and upper layer masks respectively. The mixture was then cured at 80 °C for 2 hours in a vacuum oven. The cured PDMS layers were inspected and carefully peeled off from the master mould. Furthermore, the upper PDMS layer was carefully punched to form the cell culture wells.

In the final step, the middle layer, i.e. the 200-μm deformable PDMS membrane, was fabricated and all three layers were aligned and plasma bonded to form the final cell-stretching platform. A clear degassed PDMS-crosslinker mixture (10:1 volume ratio) was spin coated at 4000 rpm for 2 mins to achieve the 200-μm thickness. Prior to plasma bonding, the deformable PDMS membrane and the lower PDMS layer with output pattern were thoroughly cleaned with isopropanol and DI water to achieve a dust-free surface. The surface of both membrane and PDMS layer was then activated using oxygen plasma (Harrick Plasma) for 45 seconds before bonding. The bonded device was then cured in a vacuum oven at 80 °C for 1 hour. The same plasma bonding procedure was repeated again to bond the prepared device with the upper PDMS layer.

2.4 Cell culture

The cell-stretching platform was sterilised with ethanol (80%) and washed with PBS for three times followed by the 20 minutes UV exposure. Furthermore, the cell-stretching platform was treated with the media and incubated for 1 hour at 37 °C and 5% CO₂ before seeding the cells to enhance biocompatibility.

Fibroblast cells were cultured in a T75 cell culture flask with DMEM/F12 (Gibco, Thermo Fisher Scientific) medium enriched with 10% FBS and 1% penicillin. The cells were maintained in a standard incubator at 37 °C and 5% CO₂ to ensure proper cell growth. For seeding the device, 70% confluent T75 flask was trypsinised with TrypLE Express

(Thermo Fisher Scientific) and resuspended into the fresh media. Aliquots of optimised seeding density of 10×10^3 Cells/150 μ l were used to seed each well of the device (i.e. total 80×10^3 Cells/1200 μ l). To ensure the cell adhesion and adequate cell growth onto the membrane, the device was incubated at 37°C with 5% CO₂ for 24 hours. Finally, following literature reporting for the fibroblast cells, predefined cyclic normal strain of 0.8% (OP1), 1.2% (OP2) and 2.2% (OP3) with 80 mbar, 0.1Hz frequency at 50% duty cycle was introduced onto the cell culture for 2 hours to obtained circumferential alignment of the fibroblast cells.^{15, 43-45} Three biological repeats were carried out with the same normal strain parameters to confirm the repeatability of the experimental results.

2.5 Cell Fixing, immunofluorescence staining and imaging

After two hours of stretching, the cells were immediately fixed with 4% PFA and then washed with PBS for three times. Further, the cells were kept in cold PBS solution at 4°C. For the quantification, cell orientation was then recorded with Nikon Eclipse Ts2 inverted phase contrast microscope using the 10x magnification. Actin and nuclei were stained respectively with ActinGreenTM 488 and NucBlueTM ReadyProbeTM Reagent (ThermoFisher Scientific) to observe the cytoskeleton reorganisation. The standard protocol was optimised and adapted to suit the developed cell stretching platform. Finally, we utilised an inverted fluorescent microscope (Olympus IX73) with 20X magnification to observe the cell nuclei and actin fibres.

3 Results and Discussion

3.1 Device characterisation

Well defined normal strain is an important aspect for achieving cell culture with predefined cellular alignment. Thus, insight into the strain experienced by the cells under investigation based on their position is important. Following, we numerically and experimentally investigated normal strain on the deformable output patterns.

Experimental approach

For experimental strain characterisation, the deformable output patterns (OP1, OP2 and OP3) were actuated by positive pressure ranging from 10 mbar to 150 mbar with a step of 10 mbar using a pressure controller. A digital camera (EO® Edmund Optics) with 1X magnification lens was placed onto the XY-stage (Zaber Technologies) and the side view along XZ-plane was recorded to observe the inflation of the output patterns. A custom made algorithm was developed for detecting the membrane inflation, fitting an ellipse into the detected curve and obtaining the displacement and normal strain parameters.

The developed algorithm was implemented in Matlab R2016a using the image processing toolbox and curve fitting

techniques. The detailed flow chart with a sample step output for the implemented algorithm is shown in Figure 4. We repeated the experiments three times for each pressure point, to obtain reliable results from the image analysis. A minimum of three image frames from each video was analysed. The averaged result for each pressure point was obtained to assure the repeatability of the data.

For computational simplicity, the ideal condition of an uniform pressure distribution on a deformed membrane with negligible thickness variation was assumed for the strain evaluation. Considering the thin-walled membrane and obtained elliptical fit of the membrane inflation, normal circumferential strain ϵ_c was calculated utilising the Hoops stress (σ_c) relationship:^{46, 47}

$$\sigma_c = \frac{Pr_i}{t} \left(1 - \frac{r_i^3}{a^2(2r_i+t)} \right) \quad (1)$$

$$\epsilon_c = \frac{\sigma_c}{E}, \quad (2)$$

where σ_c is the stress (in mbar), P is the input pressure (in mbar), r_i is the internal diameter of the inflated membrane (in mm), $t=0.2$ mm is the thickness of the membrane, a is the height along minor axis (in mm), $E = 7500$ mbar is the Young's modulus of the PDMS membrane and ϵ_c is normal circumferential strain.

Numerical approach

The numerical verification of experimental data utilised a reference 3D FEA model with OP1 ($R_{i1} = 1$ mm), OP2 ($R_{i2} = 1.5$ mm) and OP3 ($R_{i3} = 2$ mm). All material properties and boundary conditions were fixed, while the applied pressure was varied from 10 mbar to 150 mbar with a step of 10 mbar to obtain the corresponding maximum planar displacement and normal circumferential strain of all three OPs.

Figure 5(a) compares the experimental and simulation data of the maximum displacement of the membrane. As expected, all three OPs show a linear relationship between membrane displacement and the pressure. The simulation and experimental data agree well and provide an acceptable average error variance of 1.63% for OP1, 2.37% for OP2 and OP3 for 3.84%. Furthermore, the maximum applied pressure of 150 mbar results in the maximum planar displacement of 0.2 mm for OP1, 0.32 mm for OP2 and 0.53 mm for OP3. In the next step, the normal circumferential strain was compared.

Figure 5(b) shows the experimental and simulation data of the normal circumferential strain obtained with the same range of applied pressures. The data show that at 80 mbar the circumferential strains of OP1, OP2 and OP3 are $0.8 \pm 0.023\%$, $1.2 \pm 0.07\%$ and $2.2 \pm 0.081\%$ respectively.

The obtained deflection and strain data served as a guideline for the selection of the applied pressure. Considering the fibroblasts in the later experiments, we selected 80 mbar

as the input pressure for the cell-stretching platform. The next step is determining the deformation pattern of the membrane under 80 mbar, utilising the existing experimental tools and optimised FEA model. Figure 6 depicts the experimental and simulation results of the shape of the deformed ring-shaped membrane. The semi-elliptical deformation of the different OPs can be clearly observed.

As the deformation patterns are axial-symmetric, the experimentally obtained 2D curve of the deformed membrane was virtually rotated along minor axis using Matlab R2016a to obtain the 3D representation of the membrane. For comparison, the corresponding FEA model was solved. Figure 7 compares the experimental and numerical shape of the membrane of OP1 under 80 mbar. Experimental and simulation results agree well and show the maximum planar displacement (D_m) of 0.165 mm and 0.159 mm with an average error variance of 2.5% respectively. The 3D data show that the output pattern deforms into a semi-toroidal shape. The results suggest that over the central circumferential region with maximum planar displacement, cells would experience approximately the same amount of normal strain. Thus, we hypothesised that cells in this circumferential region of homogenous normal strain will experience approximately the same mechanical cues. The physiological or morphological changes observed after stretching would mainly depend on the normal strain amplitude.

3.2 Cell orientation

Our preliminary observation of the cytoskeleton reorganisation and reorientation for fibroblast cells are consistent with the previous studies and confirm the use of the developed cell stretching platform to actively manipulate cell orientation and obtain patterned cell culture with predefined cellular alignment.³⁸⁻⁴² Furthermore, the developed cell stretching platform is compatible with generic laboratory tools and protocols and ensures its adaption in pre-clinical applications.

For the quantification of cell orientation, each cell culture well was virtually divided into 8 equal quadrants of 45° each i.e. region of interests (ROI). Images from each ROI were considered for the analysis. To enhance the features and correct the background noise in phase contrast images, basic post processing steps including application of Fast Fourier Transformation (FFT) bandpass filter, sharpening and enhancing image contrast were carried out with ImageJ tools. The obtained binary image was further utilised for the analysis of cell orientation. In each ROI, orientation of the 100 cells were analysed to estimate an average orientation of the cells in each predefined quadrant. Figures 8 and 9 shows the average cell orientation for each ROI of OP1, OP2, OP3 and control.

The data and the images clearly show that the cells recognise the external mechanical cues and reorient

themselves in an optimum direction to avoid the normal strain.⁴⁸⁻⁵¹ As expected, the fibroblast cell reoriented perpendicular to the normal strain direction.^{52, 53} Moreover, the cells appeared to recognise the homogenous normal strain gradient and realign themselves along this region. Although, alignment of the cells with respect to the strain pattern was observed for all three output patterns, it was interesting to note that cell alignment along the strain profile increases with the increasing normal strain amplitude. This behaviour of fibroblasts consistent with previous reports.^{15, 37, 54}

An average cell orientation for the three output patterns (OP1, OP2 and OP3) were 131.3±14.9° for the ROI 1&2, 221±11.5° for ROI 3&4, 295.89±13.43° for ROI 5&6 and 66.82±10.41° for ROI 7&8 in Figures 8 and 9(a). In contrast, cells distributed randomly without strain and did not show any preferred orientation or significant alignment as shown in Figure 9(b).

Furthermore, we calculated the percentage cells aligned in circumferential pattern for each ROI to establish the relationship between normal strain amplitude and cell orientation. Figure 10 compares the percentage of aligned cells for the different normal strain amplitude with one-way analysis of variance (ANOVA) results. The ANOVA results suggested that all three output patterns exhibit significant cell orientation compared to the control (OP1: $p<0.001$, OP2: $p<0.001$ and OP3: $p<0.05$). However, the cell orientation of OP2 and OP3 is significantly higher ($p<0.01$) than that of OP1. Moreover, no significant difference between cell orientation was observed for the OP2 and OP3 (n.s) and clear circumferential alignment of the fibroblasts can be seen in Figure 10.

In depth molecular or biological assays for the physiological study of the stretched cells is outside the scope of the current work. Thus, we adapted the observational approach and detected noticeable cell morphology changes for all OPs to confirm any changes due to different normal strain amplitude. Interestingly, we noted that cells formed clusters under a high normal strain amplitude as observed in Figure 11.

Moreover, the fact that clear fluorescent signals were obtained from the cells after fixation for both with-strain and no-strain conditions, confirms that all the cells were firmly attached to the membrane and sustained their intracellular adhesion, Figs. 8, 9 and 11. This observation confirmed the cell viability in for our developed platform for both with-strain and no-strain conditions.

Considering the above preliminary observations, we could hypothesise that patterned cell culture can be obtained with low normal strain (0.8-1.2%) amplitude without any noticeable physiological and morphological changes. Our experimental results confirm the capability of our cell-stretching platform to actively manipulate the cell orientation to achieve patterned cell culture with circumferential cellular alignment. However,

to further confirm the capability of the platform, we modified the device to incorporate square and hexagon strain patterns and performed preliminary experiment with optimised stretching parameter (80mbar, 0.01Hz, 2Hr). As speculated, square and hexagonal shaped cellular alignment was evident from preliminary observations (Supplementary information). This observation further confirms the suitability of the developed platform to actively manipulate cellular alignment. Moreover, the developed platform can be scaled up to the standard 96 wells for clinical drug screening and for 2D mechanobiological experiments or screening of cellular response in tissue engineering and regenerative medicine. Thus, the cell-stretching platform presented here could be a critical step towards the 2D patterned cell culture for high-throughput *in-vitro* assays.

4. Conclusions

In summary, a novel cell-stretching array platform compatible with the generic tools is designed, fabricated and characterised to obtain patterned cell culture with predefined circumferential cellular alignment *in vitro*. We observed widely reported cellular realignment behaviour of the fibroblasts cells with the developed platform. The cells reoriented and realigned approximately perpendicular to the stretching direction. We also demonstrated that the cyclic normal strain can be utilised to actively manipulate the cellular alignment and achieve patterned cell culture with predefined cellular alignment. Moreover, we examined the cell reorientation with different normal strain amplitudes. Our observations suggest that the cell orientation increases with increase in normal strain amplitude. However, noticeable cell clustering was observed with higher normal strain amplitude. Although further optimisation is needed, the preliminary results allow us to hypothesise that low normal strain amplitude of 1.2% could pattern fibroblast cell culture without noticeable physiological and morphological changes.

The developed cell-stretching platform provides a new direction of bioengineering utilising cells mechanotransduction capability to actively manipulate the cellular alignment and achieve patterned cell. As shown, the developed platform can be easily optimised by redesigning the desired OPs to achieve a variety of predefined cell culture patterns, which could be critical in tissue remodelling and regenerative medicine. Furthermore, the compatibility of the developed platform with existing tools and protocols make it attractive alternative in clinical drug screening and 2D mechanobiological experiments. This device also could be very useful in modulating the local micro-environment of stem cell colonies to gain a better mechanistic insight in developmental biology. Finally, the developed cyclic cell stretching approach may also serve to understand how various mechanical cues such as strain and strain gradient influence the cell alignment and orientation which could further provide insights on cellular level interactions between cells and their immediate microenvironment.

However, further optimisation of the developed platform is needed to incorporate strain pattern with a higher resolution. Further characterisation of the platform will be needed to precisely define the minimum size and resolution of the strain pattern that can be utilised for engineering cellular alignment. Considering the advancement in Microelectro Mechanical Systems (MEMS) fabrication, we believe that the developed cell patterning approach has the potential to achieve more complex cellular micro patterns. Nonetheless, further optimisation is needed to characterise these critical parameters to enhance the capability of the developed platform and to engineer more complex cell pattern *in vitro*.

Acknowledgements

H.K and R.V. acknowledge the scholarships provided by Griffith University for their PhD studies.

References

1. B. D. Hoffman and J. C. Crocker, *Annu. Rev. Biomed. Eng.*, 2009, **11**, 259-288.
2. R. P. Brandes, *et al.*, *Antioxid. Redox Signal.*, 2014, **20**, 887-898.
3. J. H. Wang and B. P. Thampatty, *Biomech Model Mechanobiol*, 2006, **5**, 1-16.
4. C. L. Happe and A. J. Engler, *Circ. Res.*, 2016, **118**, 296-310.
5. F. H. Silver and L. M. Siperko, *Crit. Rev. Biomed. Eng.*, 2003, **31**, 255-331.
6. A. W. Orr, *et al.*, *Dev. Cell*, 2006, **10**, 11-20.
7. C. S. Chen, *J. Cell Sci.*, 2008, **121**, 3285.
8. D. J. Leong, *et al.*, in *Skeletal Biology and Medicine II: Bone and Cartilage Homeostasis and Bone Disease*, ed. M. Zaidi, Blackwell Science Publ, Oxford, 2011, vol. 1240, pp. 32-37.
9. K. Harshad, *et al.*, *Biomed. Microdevices*, 2016, **18**, 45.
10. B. D. Hoffman, *et al.*, *Nature*, 2011, **475**, 316-323.
11. S. Y. Chew, *et al.*, *Biomaterials*, 2008, **29**, 653-661.
12. S. Hoehme, *et al.*, *Proc Natl Acad Sci U S A*, 2010, **107**, 10371-10376.
13. C. C. DuFort, *et al.*, *Nat. Rev. Mol. Cell Biol.*, 2011, **12**, 308-319.
14. D. E. Ingber, *Ann Med*, 2003, **35**, 564-577.
15. Y. Cui, *et al.*, *Nat. Commun.*, 2015, **6**, 6333.
16. A. E. Ekpenyong, *et al.*, *Biochim. Biophys. Acta*, 2015, **1853**, 3105-3116.
17. H. Oda, *et al.*, *Biomaterials*, 2013, **34**, 5891-5896.
18. Y. H. Tan, *et al.*, *IEEE Trans. Biomed. Eng.*, 2010, **57**, 1816-1825.
19. S. Nishimura, *et al.*, *Prog. Biophys. Mol. Biol.*, 2008, **97**, 282-297.
20. N. J. Sniadecki, *et al.*, *Rev. Sci. Instrum.*, 2008, **79**, 044302.
21. O. Chaudhuri, *et al.*, *Nat. Methods*, 2009, **6**, 383-387.
22. S. Dhein, *et al.*, *Prog. Biophys. Mol. Biol.*, 2014, **115**, 93-102.
23. Z. He, *et al.*, *Adv. Perit. Dial.*, 2012, **28**, 2-9.
24. S. Higgins, *et al.*, *Biores Open Access.*, 2013, **2**, 212-216.
25. Q. Wang, *et al.*, *Sensors Actuators B: Chem.*, 2014, **194**, 484-491.
26. W. W. Ahmed, *et al.*, *Biomaterials*, 2010, **31**, 250-258.

27. S. P. Arold, *et al.*, *Ann. Biomed. Eng.*, 2007, **35**, 1156-1164.
28. K. Sato, *et al.*, *IJMS*, 2010, **52**, 251-256.
29. H. Kamble, *et al.*, *Sensors and Actuators A: Physical*, 2016, **252**, 174-179.
30. H. Kamble, *et al.*, *Micromachines*, 2017, **8**.
31. Y. Huang and N. T. Nguyen, *Biomed. Microdevices*, 2013, **15**, 1043-1054.
32. Y. Kamotani, *et al.*, *Biomaterials*, 2008, **29**, 2646-2655.
33. Y. J. Chang, *et al.*, *Nanomedicine*, 2013, **9**, 345-355.
34. Y. Iwadate and S. Yumura, *Biotechniques*, 2009, **47**, 757-767.
35. I. Guido, *et al.*, *Microelectron. Eng.*, 2012, **97**, 379-382.
36. Z. L. Shen, *et al.*, *Biophys. J.*, 2008, **95**, 3956-3963.
37. L. Wang, *et al.*, *ACS applied materials & interfaces*, 2015, **7**, 15088-15097.
38. S. Chagnon-Lessard, *et al.*, *Integr. Biol. (Camb.)*, 2017, DOI: 10.1039/c7ib00019g.
39. P. Feng, *et al.*, *Biomedical engineering online*, 2016, **15**, 63.
40. Y. Chen, *et al.*, *Proc Natl Acad Sci U S A*, 2013, **110**, E2352-2361.
41. S. C. Tyagi, *et al.*, *J Cell Physiol*, 1998, **176**, 374-382.
42. T. W. Gilbert, *et al.*, *Tissue Eng.*, 2007, **13**, 1313-1323.
43. H.-S. Yu, *et al.*, *J. Tissue Eng.*, 2015, **6**, 2041731415618342.
44. E. H. Kim, *et al.*, *BioChip Journal*, 2015, **9**, 306-312.
45. F. Boccafroschi, *et al.*, *Cell adhesion & migration*, 2007, **1**, 124-128.
46. J. C. Wang, 2005.
47. P. A. McHALE and J. C. Greenfield, *Circ. Res.*, 1973, **33**, 303-312.
48. H. Abiko, *et al.*, *J Cell Sci*, 2015, **128**, 1683-1695.
49. U. Faust, *et al.*, *PLoS One*, 2011, **6**, e28963.
50. C. Tamiello, *et al.*, *Sci. Rep.*, 2015, **5**, 8752.
51. J. Qian, *et al.*, *PLoS One*, 2013, **8**, e65864.
52. Y. Morita, *et al.*, *Cell Biochem. Biophys.*, 2013, **67**, 1249-1259.
53. A. Pavesi, *et al.*, *Sci. Rep.*, 2015, **5**, 11800.
54. S. Chagnon-Lessard, *et al.*, *Integr. Biol. (Camb.)*, 2017, **9**, 607-618.

Fig. 1. The pneumatically actuated cell-stretching platform: (a) Schematic illustration of the cell stretching platform in the OFF state. (b) Schematic representation of working principle of the device with OP1 and OP3 in the ON state. (c) FEA model results illustrating the OP1 and OP3 behaviour during ON state. (d) Lower PDMS layer (4x2) with deformable output patterns and control pattern. (e) Actual multilayer PDMS cell-stretching platform.

Fig.2 Simulation results of the maximum planar displacement versus the inner post radius (R_i) versus (Inset: optimised geometry for the deformable OP1, OP2 and OP3, and the membrane deformation results for OPs with applied pressure of 80 mbar).

Fig.3 Fabrication process for cell stretching platform.

Fig.4. Developed algorithm to detect the inflated membrane to fit the curve and calculates the displacement and normal strain parameters. (a) Input: Reference Image and scale (pixels per mm). (b) Image enhancement process: Normalised contrast and sharpening. (c) Thresholding and Edge detection using canny. (d) Dilation operation and noise removal for curve detection. (e) Fill the holes to enhance mask calculations by eliminating minor curves. (f) Calculate mask and get minimum Y coordinates along the curve for selection. (g) Generate symmetric reflected curve along major axis for curve fitting. (h) Apply least Square criterion for curve fitting. (i) Calculate curve parameters and display results.

Fig.5 Comparison of the simulation and experimental data: (a) Maximum displacement versus pressure. (Inset: simulation results with pressure of 150 mbar); (b) Circumferential strain versus pressure. (Inset: experimental setup).

Fig.6 Comparison of the membrane deformation for OP1, OP2 and OP3 with pressure input 80 mabr. (a) Experimental results (actual membrane deformation in XZ plane) (b) Simulation results (FEA model results for membrane deformation in XZ plane) .

Fig.7 Output membrane deformation pattern of OP1: (a) Experimental results. (b) Simulation results.

Fig. 8 Averaged cell orientation in degrees over the ROI for OP1 and OP2 with corresponding 10X cell images of each ROI and the COMSOL strain distribution map. Scale bar is 100 μm .

Fig. 9 Averaged cell orientation in degrees over the ROI for OP3 (a) and Control (b) with corresponding 10X cell images of each ROI and the COMSOL strain distribution map. Scale bar is 100 μm .

Fig. 10 Percentage of cells in given direction over the ROI and ANOVA analysis for OP1, OP2, OP3 and Control. The cell orientation varies according to normal strain amplitude (** $p < 0.001$, and * $p < 0.05$ versus control; ## $p < 0.01$ versus the lowest normal strain amplitude; n.s., not significant). Results are expressed as means $\pm$ s.d. ($n = 3$) and data were analysed using one-way ANOVA followed by post hoc Bonferroni's test.

Fig. 11 Cytoskeleton organisation for OP1 (a), OP2 (b) and OP3 (c) and Control (d) with 20x Objective (scale bar: 50 μm) after 2 hours.

We report a pneumatically actuated cell-stretching well array to engineer cell pattern in-vitro.