Achilles and patellar tendinopathy display opposite changes in elasticity: An ultrasound shear wave elastography study.

Authors:

Brooke Kaye Coombes (PhD)¹; Kylie Tucker¹ (PhD); Bill Vicenzino (PhD)²; Viana Vuvan²; Rebecca Mellor²; Luke Heales (PhD)³; Antoine Nordez (PhD)⁴; François Hug (PhD)^{2,4,5}

Author affiliations:

¹School of Biomedical Sciences, The University of Queensland, Brisbane Australia

²School of Health and Rehabilitation Sciences, The University of Queensland, Brisbane, Australia

³School of Human, Health and Social Science, Division of Physiotherapy, Central Queensland University, Rockhampton, Australia

⁴Faculty of Sport Sciences, Laboratory “Movement, Interactions, Performance” (EA 4334), University of Nantes, Nantes, France

⁵Institut Universitaire de France (IUF), Paris, France.

Corresponding author: Dr Brooke Coombes

University of Queensland, School of Biomedical Sciences, Otto Hirschfeld Bld 81, St Lucia 4072

P: +614 3750 4508 E: b.coombes@uq.edu.au

Running Head: Tendinopathy: A shearwave elastography study

Acknowledgements: The authors thank Wolbert Van den Hoorn and Ilze Willis for assistance with Matlab.

25 **Funding source:** Dr Coombes is in receipt of a University of Queensland Postdoctoral
26 Fellowship for Women and received a travel fellowship by the France-Australia Science
27 Innovation Collaboration. Support was received from The French Ministry of Sport (17-
28 R-04) for publication costs.

29 **Conflict of Interest:** No conflict of interest, financial or otherwise, are declared by the
30 authors.

31

32 **Original research**

33 **Word count 2868**

34

ABSTRACT

Purpose: To compare tendon mechanical properties (elastic modulus) and structural properties (thickness) of healthy individuals with those with Achilles or patellar tendinopathy.

Methods: Sixty-seven participants (22 Achilles tendinopathy, 17 patellar tendinopathy, 28 healthy controls) were recruited between March 2015-March 2016. Shear wave velocity (SWV), an index of tissue elasticity, and tendon thickness were measured bilaterally at mid-tendon and insertional regions of Achilles and patellar tendons by an examiner blinded to group. Analysis of covariance, adjusted for age, body mass index and sex was used to compare differences in tendon thickness and SWV between the two tendinopathy groups (relative to controls) and regions. Tendon thickness was included as a covariate for analysis of SWV.

Results: Compared to controls, participants with Achilles tendinopathy had lower SWV at the distal insertion (Mean difference MD; 95% CI: -1.56; -2.49 to -0.62 m/s; $P<0.001$), and greater thickness at the mid-tendon (MD 0.19; 0.05 to 0.33 cm; $P=0.007$). Compared to controls, participants with patellar tendinopathy had higher SWV at both regions (MD 1.25; 0.40 to 2.10 m/s; $P=0.005$), and greater thickness proximally (MD 0.17; 0.06 to 0.29 cm; $P=0.003$).

Conclusion: Compared to controls, participants with Achilles and patellar tendinopathy displayed lower Achilles tendon elasticity and higher patellar tendon elasticity respectively. More research is needed to explore whether maturation, aging or chronic load underlie these findings and whether current management programs for Achilles and patellar tendinopathy need to be tailored to the tendon.

Key words: jumper's knee; Aixplorer; ultrasound, elasticity, stiffness, musculoskeletal,

60 rehabilitation, pain

61

INTRODUCTION

Painful disorders of the Achilles and patellar tendons are a major problem in competitive and recreational sports as well as the sedentary population. More than half (52%) of elite runners develop Achilles tendinopathy during their lifetime,¹ while the prevalence of current or previous patellar tendinopathy in volleyball or basketball athletes is similar (50-55%).² Ultrasonography is widely used for diagnosis of tendinopathy, however classical ultrasound findings show limited correlation with symptoms^{3, 4} and lack fundamental information on tissue mechanical properties.

Shear wave elastography is a relatively new ultrasound imaging technique that allows non-invasive estimation of soft-tissue viscoelastic properties in vivo by measurement of Shear Wave Velocity (SWV) (in meters per second) generated by the ultrasound pulse.⁵ Importantly, as the SWV may be determined from a relatively small region of tissue (as opposed to the whole structure) it can provide direct measurement of specific regions of interest within tendon,^{6, 7} which is important because discrete areas of pathology are reported in people with tendinopathy.⁸

Using this technique, lower elasticity was demonstrated in patients with Achilles tendinopathy,^{9, 10} whereas both higher¹¹ and lower¹⁰ elasticity were reported in separate studies of patients with patellar tendinopathy. By estimating patellar tendon stiffness from the relationship between the force applied to the tendon and its associated proximal insertion displacement assessed through B-mode ultrasound, inconsistent findings are also reported (no change¹² or decreased⁸ stiffness). However, this indirect method to assess tendon stiffness has two major drawbacks: i) the tendon force is indirectly estimated (as the force is applied externally to a body segment, rather than directly on the tendon), and ii) only global

properties of the whole tendon can be estimated. This is surprising given the similar histological appearance in patellar and Achilles tendinopathy,¹³ but other factors, including age, body composition and training might also modulate the mechanical properties of tendon.¹⁴ For example, athletes with patellar tendinopathy are significantly younger (mean age 23 ± 3 years), taller (mean height 185 ± 10 cm) and heavier (mean weight 77 ± 11 kg), although of similar body mass index (BMI) (mean BMI 23 ± 2 kg/m²) than those without symptoms.¹⁵ In comparison, participants with Achilles tendinopathy are older (mean age 53 ± 12 years) and have higher BMI (mean 35 ± 8 kg/m²) than control participants.¹⁶

Methodological differences surrounding measurement of tendon properties may also contribute to the variability between studies.¹⁷ Of relevance to the shear wave elastography technique for tendons, shear wavelengths can be greater than tendon thickness, leading to guided wave propagation.¹⁸ Within this context, the relationship between group velocity of the shear wave estimated by commercialized ultrasound shear wave elastography techniques and tendon elasticity is affected by tendon thickness. Thus, information about both SWV and thickness must be sought to interpret changes in SWV as changes in elasticity. In addition, some authors⁹ have reported measurement artefacts during shear wave elastography acquisition, including signal void areas within an elastogram. While their presence has been suggested to be a potential marker of intratendinous damage,⁹ comprehensive evaluation of its presence is lacking.

The purpose of our study was to determine whether tendon elasticity and thickness of healthy individuals differ from those with Achilles or patellar tendinopathy using shear wave elastography. We proposed that regional differences in SWV and thickness would be present in those with tendinopathy compared to controls. As other factors may also modulate the

elastic properties of connective tissues, we also investigated the relationship between SWV and age, sex and BMI. Last, we aimed to quantify elastography signal void in tendinopathic and healthy tendon.

MATERIALS AND METHODS

Participants with or without tendinopathy were consecutively recruited by social media and local advertisement between March 2015 and March 2016. Inquiries were directed to an online screening questionnaire and eligible participants were invited to attend physical screening by a musculoskeletal physiotherapist not involved in testing. The inclusion criteria for patients with tendinopathy were self-reported tendon pain over the Achilles or patellar tendon of at least 20 on a 0-100 numerical rating scale (NRS), and of duration 12 weeks or more. Physical examination confirmed tendon pain on palpation and loading (by calf raise or decline squat). The number of single leg calf raises or decline squats performed before pain onset was recorded. Exclusion criteria were corticosteroid injection (in previous 6 months), local surgery or other specific pathology (e.g., fracture, diabetes, inflammatory, systemic or neurological disease, malignancy, radiculopathy). All participants were excluded if they had experienced back or other leg symptoms requiring treatment or limiting work/sport in the previous six months. This study was approved by the University of Queensland Medical Research Ethics Committee and all participants provided written informed consent.

Participants with tendinopathy completed either the Achilles¹⁹ or Patellar²⁰ versions of the Victorian Institute of Sport Assessment (VISA), to provide validated information about pain and disability for their (most) affected leg. Total scores range from 0 to 100, 100 representing an asymptomatic, fully performing individual. In addition to basic demographic information (age, sex, height and weight), all participants completed the Active Australia Questionnaire,

to measure physical activity during the preceding week.²¹ Total activity time (minutes) was calculated by adding the time spent in walking and moderate activity and twice the time spent in vigorous activity.

Elastography acquisition Upon arrival participants rested in lying for five minutes prior to examination using the supersonic shear imaging technique (Aixplorer version 8.2, Supersonic Imagine, Aix-en-Provence, France) with 50 mm linear transducer (15-4MHz). A single operator (BKC) performed all scans to eliminate inter-observer variability.²² Interday reliability was examined by testing both legs in six healthy participants on two occasions separated by 48 hours. Intraclass correlation coefficients (ICC) ranged from 0.71 to 0.80 for measurement of SWV and 0.78 to 0.97 for thickness measurements (Table 1).

Examination was performed blind to the physical examination and group by scanning bilateral patellar and Achilles tendons in a randomised order for all participants. For Achilles measurements, participants lay in prone with their leg extended and ankle resting over the edge of the plinth. The participant's resting ankle angle was measured using a hand-held goniometer. For patellar tendon measurement, participants lay in supine, supported in 30° knee flexion and neutral hip rotation. Based on previous report of the slack angle of the Achilles tendon ($43.7 \pm 3.2^\circ$ plantarflexion)²³ and unpublished observations of the patellar tendon, both positions were assumed to have some passive tension on the respective tendons. The amount of signal void within each ROI, defined as the percentage of all pixels without elastography colour was also calculated. For SWV measurement, we made sure that regions with signal void were not confused with a low value. Participants were told to remain completely relaxed during testing.

The ultrasound transducer head was aligned in the longitudinal plane with collagen fibres, applying minimal pressure.²⁴ Tendon thickness was measured from longitudinal B-mode images using the inbuilt Aixplorer distance function. For the Achilles tendon, the maximum perpendicular distance between anterior and posterior tendon boundaries was measured over the free tendon (2-4 cm proximal to the Achilles insertion) and immediately proximal to the calcaneal attachment. For the patellar tendon, the maximum perpendicular distance immediately distal to the apex of the patellar and at the midpoint between patellar and tibial attachments was measured. Using the shear wave elastography mode (penetration; persistence low; smoothing 5), the transducer was held for ~10 s duration, and two or three measurements were repeated in each location. For the Achilles tendon, separate images were collected for measurement of mid-tendon and insertional regions, while both patellar regions could be visualised without moving the transducer.

Data extraction Elastography clips (10 s duration, sampling rate of 1.5-1.8 Hz) were converted into PNG images (~20-26 images), then processed offline using customised Matlab scripts (R2016a, The Mathworks, Natick, MA, USA) by a research assistant who was blinded to group. The following regions of interest (ROI) were manually traced on the adjacent B-mode image keeping within the peritendinous boundary: For the Achilles tendon, ROI were measured at the free tendon (ROI length ~3 cm) and immediately proximal to the calcaneal attachment (ROI length ~1 cm). For the patellar tendon, ROI were measured mid-way between proximal and distal attachments (ROI length ~2 cm) and 0.5 cm from the apex of the patella (ROI length ~0.5 cm). The mean SWV (m/s) was averaged over the ROIs of consecutive images. The amount of signal void within each ROI, defined as the percentage of all pixels without elastography colour, was also calculated. For SWV measurement, the Matlab script ensured that regions with signal void were not confused with a low value.

Examples of Achilles and patellar elastograms (highlighting signal void) are provided in Figure 1 and Figure 2, respectively.

Statistical analysis For Achilles tendinopathy (AT) and patellar tendinopathy (PT) groups, only the affected leg or self-reported worst leg (in bilateral cases) was included in analysis. For healthy control (HC) participants, we randomly matched Achilles and patellar tendons to ensure an equivalent proportion of left/right legs in the analysed sample, as recommended by previous studies.²⁵ Repeated measures ANCOVAs were used to compare tendon SWV and thickness between groups (AT vs HC or PT vs HC) and regions (mid-tendon vs insertion). Where significant group-by-region interaction was observed, generalised linear models (GLM) were used to estimate mean differences (MD) and 95% confidence intervals (CI) between groups for each tendon region. Age, BMI and sex were included as covariates in all models. In addition, we included tendon thickness as a covariate in analyses of SWV, based on evidence that tendon thickness may affect SWV.^{18, 26} We dichotomised signal void as present if it affected more than 5% of pixels within the ROI and differences between groups were compared using Pearson Chi-square statistic.

Receiver operating characteristic (ROC) analysis was performed to determine the threshold, specificity and sensitivity to differentiate HC from AT or PT, using the Youden's index to identify the cut-point.²⁷ Finally, Pearson correlations were computed to identify relationships between SWV and demographic or injury factors (the latter for tendinopathy groups only). Statistical analysis was performed using SPSS (22.0 IBM Corp). Significance was set at $P < 0.05$.

RESULTS

Participants Sixty-seven adults (22 with AT, 17 with PT and 28 HC who were free of tendon pain) were recruited. Demographic and injury information is summarised in Table 2 and reasons for participant exclusion are listed in Supplemental Figure 1. Participants with AT were significantly older than those with PT ($P<0.001$) and had higher BMI than HC ($P<0.001$). Total activity time was not statistically different between the three groups ($P=0.343$). Similar symptom duration ($P=0.943$) and pain/disability levels ($P=0.781$) were found for AT and PT groups. The resting ankle angle adopted during Achilles tendon testing was not different between groups (mean plantarflexion angle $24.3\pm 5.0^\circ$).

Achilles tendinopathy Compared to HC, participants with AT had greater thickness at the mid-Achilles (MD 95% CI: 0.19; 0.05 to 0.33 cm; $P=0.007$), but not at the Achilles insertion ($P=0.80$) (group-by-region interaction, $P=0.035$, Figure 3a). Compared to HC, participants with AT had lower SWV at the Achilles insertion (MD -1.56; -2.49 to -0.62 m/s; $P<0.001$), but not mid-tendon region ($P=0.456$) (group-by-region interaction, $P<0.001$, Figure 3c). Signal void of more than 5% of the ROI was present in four HC (8.9%) and 10 AT (45.5%) cases (Pearson Chi-square $P=0.001$).

Using ROC analysis, mid-Achilles thickness above 0.64 cm, SWV below 9.7 m/s and signal void above 1.5% of the ROI were identified as cut-points to differentiate HC and AT (Table 3). Lower Achilles SWV was associated with higher age ($r=-0.49$, $P<0.001$), greater BMI ($r=-0.53$, $p<0.001$), greater pain and disability as measured by VISA-A ($r=0.49$, $P=0.046$) and fewer single leg calf raises before pain onset ($r=0.646$, $P=0.001$).

Patellar tendinopathy Compared to HC, participants with PT had greater thickness at the proximal patellar (MD 0.17; 0.06 to 0.29 cm; $P=0.003$), but not mid-patellar ($P=0.31$) region

(group-by-region interaction, $P=0.01$, Figure 3b). Compared to HC, participants with PT had higher SWV of the patellar tendon at both regions (MD 1.25; 0.40 to 2.10 m/s; $P=0.005$, Figure 3d). Signal void of more than 5% of the ROI was present in 0 HC and 3 PT (17%) cases (Pearson Chi-square $P=0.021$).

Using ROC analysis, proximal patellar thickness above 0.62 cm, SWV above 6.9 m/s and signal void above 0.7% of the ROI were identified cut-points to differentiate HC and PT (Table 3). Lower proximal patellar SWV was correlated with age ($r=-0.368$, $P=0.013$) only.

DISCUSSION

Using the non-invasive ultrasound-based technique of shear wave elastography, we observed lower insertional Achilles tendon elasticity and higher patellar tendon elasticity in participants with tendinopathy relative to healthy controls. Lower elasticity in Achilles tendinopathy is consistent with two recent studies using shear wave elastography,^{9, 10} and other force-deformation methods.²⁸ However, evidence for changes in patellar tendon mechanical properties in tendinopathy varies between methods^{8, 29} and studies.^{10, 11} Using shear wave elastography, Zhang (2014) found significantly greater elasticity and thickness for the proximal patellar tendon in 13 athletes with unilateral PT, compared to both the unaffected tendon and healthy control participants.¹¹ In contrast, Dirrichs et al (2016) reported lower elasticity in symptomatic compared to asymptomatic tendons in 38 participants with patellar tendinopathy.¹⁰ However, Dirrichs et al (2016) reported using a much smaller ROI (standardised size of 1 mm) placed in the stiffest area of tendon.¹⁰

Although causal relationships cannot be inferred from this cross-sectional study, it is possible that different pathobiological processes associated with maturation and aging may be at play

in patellar and Achilles tendinopathy, respectively. Animal work demonstrates that maturation and aging have different influences on the physical, chemical and mechanical properties of tendon.³⁰ In their experimental study in rats, Vogel et al (1980) observed a sharp rise in elasticity during maturation and a smaller but significant decrease during aging.³⁰ These changes in mechanical parameters were closely related to cross-linking of collagen. Significant increases in tendon elasticity with maturation³¹ may explain why adolescent athletes have three times greater odds of patellar than Achilles tendinopathy.³² In our study, older age was associated with lower Achilles tendon elasticity, which is in agreement with previous investigations using shear wave elastography,^{7, 24} and other methods.³³ Differences in chronic load may alternatively contribute to our findings. Cassel et al (2015) reported that athletes with patellar tendinopathy were taller and heavier and performed higher training volumes than athletes without patellar tendinopathy,³² attributes which may induce adaptation by increased tendon elasticity. Conversely, sedentary lifestyles adopted by some people with Achilles tendinopathy may contribute to lower elasticity. Interestingly, biomechanical studies demonstrate that the effects of stress deprivation on connective tissue are more prolonged at the insertion than mid-tissue region,³⁴ potentially explaining the significant reduction of mechanical properties at the Achilles insertion. Of clinical importance, Achilles tendon elasticity was significantly correlated with self-reported pain and disability over the preceding week and with a rudimentary measure of the loading capacity of the Achilles tendon.

The strengths of this study include the blinded design and concurrent investigation of two different lower limb tendinopathies. We acknowledge that interpretation of differences in SWV as differences in elasticity is not straightforward, as larger tendon thickness can lead to a higher SWV due to the effects of guided wave propagation, independent of any change in

actual tissue elasticity.^{18, 35} In the present study, no between-group difference in thickness was observed for the mid-patellar tendon and the Achilles tendon insertion, increasing our confidence in our interpretations that mechanical properties were altered. However, it is possible the technique failed to detect a less stiff (but thicker) mid-Achilles tendon. This study extends upon previous ones by quantifying signal void within the measured ROI. Signal void was highly specific (82-96%) to cases with tendinopathy. Although currently speculative, its presence may be due to intra-tendinous swelling or tears,⁹ greater attenuation of shear waves as a result of altered viscosity,²⁶ or severe matrix disruption and increased anisotropy.³⁶ Further studies are needed to determine the impact of attenuation on SWV in the considered region, and relate signal void to changes in matrix proteoglycan concentration due to tendinopathy.

Limitations of the current study include the moderate reliability estimates, small sample size and heterogenous tendinopathy population, which included participants with symptoms at either mid-tendon or insertional regions. We chose to recruit both male and female control participants with a range of ages, rather than individually match participants on the basis of age or sex. Although not without limitation, statistical adjustment by including age, sex and BMI as covariates in all models, was performed in an attempt to explore which factors modulate changes in tendon properties. Other potential confounding factors such as sporting history and muscle strength were not measured.

In summary, in the positions tested here, Achilles tendon SWV less than 9.7 m/s and patellar tendon SWV more than 6.9 m/s showed high specificity (81% and 82% respectively) and sensitivity (79% and 77%) for detecting Achilles and patellar tendinopathy respectively,

311 suggesting this method may have important clinical utility for early detection of patients at
312 risk of tendon pathology.

313

314 PERSPECTIVE

315 Achilles and patellar tendinopathy share similar histopathology, but affect different age
316 groups. The present study suggests that the two lower limb tendinopathies display different
317 mechanical properties when compared to tendons from healthy individuals. More research is
318 needed to explore whether maturation, aging or chronic load underlie these findings and
319 whether current management programs for Achilles and patellar tendinopathy need to be
320 tailored to the tendon.

Table 1: Inter-day reliability (48 hour interval) for measurement of tendon thickness and shear wave velocity in six healthy individuals. ICC: Intraclass correlation coefficient.

Tendon region	Session 1	Session 2	ICC (95%CI)
	mean (SD)	mean (SD)	
	Thickness (cm)		
Mid-Achilles	0.54 (0.22)	0.57 (0.22)	0.97 (0.92, 0.99)
Insertional Achilles	0.40 (0.08)	0.41 (0.06)	0.78 (0.47, 0.92)
Mid-patellar	0.34 (0.05)	0.34 (0.06)	0.84 (0.59, 0.94)
Proximal patellar	0.53 (0.07)	0.53 (0.07)	0.82 (0.56, 0.94)
	Shear wave velocity (m/s)		
Mid-Achilles	10.4 (1.1)	10.8 (0.8)	0.71 (0.24, 0.91)
Insertional Achilles	9.8 (1.6)	10.0 (1.2)	0.69 (0.11, 0.92)
Mid-patellar	6.1 (0.8)	6.0 (0.7)	0.71 (0.22, 0.91)
Proximal patellar	6.4 (1.2)	6.4 (1.2)	0.80 (0.42, 0.94)

Table 2: Demographic and injury characteristics for Healthy Control, Patellar tendinopathy and Achilles tendinopathy cohorts.

	Healthy Control	Patellar Tendinopathy	Achilles Tendinopathy
n	28	17	22
<i>Demographic information</i>			
Age (years)	38.3 (16.7)	29.4 (6.6)	47.5 (11.4)*
BMI (kg/m ²)	22.8 (2.7)	25.4 (3.3)	28.1 (5.3)*
Female sex n (%)	17 (61%)	4 (24%)*	9 (41%)
Total activity time (mins)	582.2 (451.9)	797.9 (481.3)	642.4 (491.3)
<i>Injury information</i>			
Injury duration (years)		3.8 (3.4)	3.7 (6.9)
VISA-A or VISA-P (0-100)		57.1 (14.4)	58.6 (16.5)
Worst pain (NRS, 0-10)		5.2 (2.3)	6.0 (2.1)
Average pain (NRS, 0-10)		3.4 (2.1)	4.2 (1.9)
Resting pain (NRS, 0-10)		1.2 (2.4)	1.4 (1.5)
Calf raises (repetitions to pain onset)		NA	9.7 (8.6)
Decline squats (repetitions to pain onset)		2.1 (1.7)	NA
<i>Symptom distribution n (%)</i>			
Bilateral tendinopathy		9 (53%)	14 (64%)
Proximal: distal symptoms		15 (88%): 2 (12%)	NA
Mid-tendon: insertional symptoms		NA	19 (87%): 9 (41%)

Data represents mean (SD) unless otherwise specified. * Significant (p<0.05) differences compared to healthy controls.

Table 3: Results of Receiver Operating Characteristic analysis.

	Mean Threshold	Specificit y	Sensitivit y
Thickness			
Mid-Achilles	≥ 0.64 cm	89.3 %	72.7 %
Insertional Achilles	≥ 0.44 cm	67.9 %	59.1 %
Proximal patellar	≥ 0.62 cm	100.0 %	58.8 %
Mid-Patellar	≥ 0.33 cm	50.0 %	82.4 %
Shear wave velocity			
Mid-Achilles	≤ 6.80 m/s	68.2 %	75.0 %
Insertional Achilles	≤ 9.68 m/s	81.0 %	78.6 %
Proximal patellar	≥ 6.90 m/s	82.1 %	76.5 %
Mid-Patellar	≥ 7.03 m/s	92.9 %	35.3 %
Signal void			
Mid-Achilles	≥ 1.5 %	82.1 %	59.1 %
Insertional Achilles	≥ 5.1 %	96.4 %	38.1 %
Proximal patellar	≥ 0.7 %	96.4 %	41.2 %
Mid-Patellar	≥ 0.5 %	82.1 %	41.2 %

333

334 Thresholds for tendon thickness (in cm), shear wave velocity (in m/s) and signal void (as a
 335 percentage of total ROI) to differentiate healthy and tendinopathic tendon have been
 336 calculated using the Youden's index. Specificity and sensitivity were generated at these cut-
 337 points.

338

Figure 1: Example shear wave elastograms of the Achilles tendon from healthy control (1a) and Achilles tendinopathy participants (1b and 1c). Signal void highlighted (asterisk) in 1c.

Figure 2: Shear wave elastograms of the patellar tendon from healthy control (1a) and Achilles tendinopathy participants (1b and 1c). Signal void highlighted (asterisk) in 2c.

Figure 3: Thickness and Shear wave velocity (SWV) of the Achilles tendon (a and c) and patellar tendons (b and d) for Healthy Control (Blue), Achilles tendinopathy (Red) and patellar tendinopathy (orange). Values represent mean and SD. *Significant ($P < 0.05$) differences between tendinopathy and healthy control groups (determined by Generalised linear model adjusted for age, body mass index, sex and thickness (for SWV measures).

352 Supplemental Content Figure 1: Recruitment numbers and participant eligibility information

353

354

355

356 References:

- 357 1. Kujala UM, Sarna S, Kaprio J (2005) Cumulative incidence of achilles tendon
358 rupture and tendinopathy in male former elite athletes. Clin J Sport Med.15:133
- 359 2. Lian OB, Engebretsen L, Bahr R (2005) Prevalence of jumper's knee among
360 elite athletes from different sports: a cross-sectional study. Am J Sports Med.33:561
- 361 3. Cook JL, Khan KM, Kiss ZS, Coleman BD, Griffiths L (2001) Asymptomatic
362 hypoechoic regions on patellar tendon ultrasound: A 4-year clinical and ultrasound
363 followup of 46 tendons. Scand J Med Sci Sports.11:321
- 364 4. Emerson C, Morrissey D, Perry M, Jalan R (2010) Ultrasonographically
365 detected changes in Achilles tendons and self reported symptoms in elite gymnasts
366 compared with controls--an observational study. Man Ther.15:37
- 367 5. Bercoff J, Tanter M, Fink M (2004) Supersonic shear imaging: a new
368 technique for soft tissue elasticity mapping. IEEE transactions on ultrasonics,
369 ferroelectrics, and frequency control.51:396
- 370 6. Dewall RJ, Jiang J, Wilson JJ, Lee KS (2014) Visualizing tendon elasticity in
371 an ex vivo partial tear model. Ultrasound Med Biol.40:158
- 372 7. Slane LC, Martin J, DeWall R, Thelen D, Lee K (2016) Quantitative ultrasound
373 mapping of regional variations in shear wave speeds of the aging Achilles tendon.
374 Eur Radiol.
- 375 8. Helland C, Bojsen-Moller J, Raastad T, Seynnes OR, Moltubakk MM,
376 Jakobsen V, Visnes H, Bahr R (2013) Mechanical properties of the patellar tendon in
377 elite volleyball players with and without patellar tendinopathy. Br J Sports
378 Med.47:862

- 379 9. Aubry S, Nueffer JP, Tanter M, Becce F, Vidal C, Michel F (2015)
380 Viscoelasticity in Achilles tendonopathy: quantitative assessment by using real-time
381 shear-wave elastography. *Radiology*.274:821
- 382 10. Dirrichs T, Quack V, Gatz M, Tingart M, Kuhl CK, Schradling S (2016) Shear
383 Wave Elastography (SWE) for the Evaluation of Patients with Tendinopathies. *Acad*
384 *Radiol*.
- 385 11. Zhang ZJ, Ng GY, Lee WC, Fu SN (2014) Changes in morphological and
386 elastic properties of patellar tendon in athletes with unilateral patellar tendinopathy
387 and their relationships with pain and functional disability. *PLoS ONE*.9:e108337
- 388 12. Kongsgaard M, Qvortrup K, Larsen J, Aagaard P, Doessing S, Hansen P,
389 Kjaer M, Magnusson SP (2010) Fibril morphology and tendon mechanical properties
390 in patellar tendinopathy: effects of heavy slow resistance training. *Am J Sports*
391 *Med*.38:749
- 392 13. Maffulli N, Testa V, Capasso G, Ewen SW, Sullo A, Benazzo F, King JB
393 (2004) Similar histopathological picture in males with Achilles and patellar
394 tendinopathy. *Med Sci Sports Exerc*.36:1470
- 395 14. Bohm S, Mersmann F, Arampatzis A (2015) Human tendon adaptation in
396 response to mechanical loading: A systematic review and meta-analysis of exercise
397 intervention studies on healthy adults. *Sports Med*.1
- 398 15. Zwerver J, Bredeweg SW, van den Akker-Scheek I (2011) Prevalence of
399 Jumper's knee among nonelite athletes from different sports: a cross-sectional
400 survey. *Am J Sports Med*.39:1984
- 401 16. Scott RT, Hyer CF, Granata A (2013) The correlation of Achilles tendinopathy
402 and body mass index. *Foot Ankle Spec*.6:283

- 403 17. Seynnes OR, Bojsen-Moller J, Albracht K, Arndt A, Cronin NJ, Finni T,
404 Magnusson SP (2015) Ultrasound-based testing of tendon mechanical properties: a
405 critical evaluation. *Journal of applied physiology* (Bethesda, Md : 1985).118:133
- 406 18. Helfenstein-Didier C, Andrade RJ, Brum J, Hug F, Tanter M, Nordez A,
407 Gennisson JL (2016) In vivo quantification of the shear modulus of the human
408 Achilles tendon during passive loading using shear wave dispersion analysis. *Phys*
409 *Med Biol*.61:2485
- 410 19. Robinson JM, Cook JL, Purdam C, Visentini PJ, Ross J, Maffulli N, Taunton
411 JE, Khan KM, Victorian Institute Of Sport Tendon Study G (2001) The VISA-A
412 questionnaire: a valid and reliable index of the clinical severity of Achilles
413 tendinopathy. *Br J Sports Med*.35:335
- 414 20. Visentini PJ, Khan KM, Cook JL, Kiss ZS, Harcourt PR, Wark JD (1998) The
415 VISA score: an index of severity of symptoms in patients with jumper's knee (patellar
416 tendinosis). *Victorian Institute of Sport Tendon Study Group. J Sci Med Sport*.1:22
- 417 21. Australian Institute of Health and Welfare. The Active Australia Survey: A
418 guide and manual for implementation, analysis and reporting. Canberra2003.
- 419 22. Peltz CD, Haladik JA, Divine G, Siegal D, van Holsbeeck M, Bey MJ (2013)
420 ShearWave elastography: repeatability for measurement of tendon stiffness. *Skeletal*
421 *Radiol*.42:1151
- 422 23. Hug F, Lacourpaille L, Maisetti O, Nordez A (2013) Slack length of
423 gastrocnemius medialis and Achilles tendon occurs at different ankle angles. *J*
424 *Biomech*.46:2534
- 425 24. Aubry S, Risson JR, Kastler A, Barbier-Brion B, Siliman G, Runge M, Kastler
426 B (2013) Biomechanical properties of the calcaneal tendon in vivo assessed by
427 transient shear wave elastography. *Skeletal Radiol*.42:1143

- 428 25. Bohm S, Mersmann F, Marzilger R, Schroll A, Arampatzis A (2015)
429 Asymmetry of Achilles tendon mechanical and morphological properties between
430 both legs. *Scand J Med Sci Sports*.25:e124
- 431 26. Brum J, Bernal M, Gennisson JL, Tanter M (2014) In vivo evaluation of the
432 elastic anisotropy of the human Achilles tendon using shear wave dispersion
433 analysis. *Phys Med Biol*.59:505
- 434 27. Akobeng AK (2007) Understanding diagnostic tests 3: Receiver operating
435 characteristic curves. *Acta Paediatr*.96:644
- 436 28. Arya S, Kulig K (2010) Tendinopathy alters mechanical and material
437 properties of the Achilles tendon. *Journal of applied physiology (Bethesda, Md :*
438 *1985)*.108:670
- 439 29. Coupe C, Kongsgaard M, Aagaard P, Vinther A, Boesen M, Kjaer M,
440 Magnusson SP (2013) Differences in tendon properties in elite badminton players
441 with or without patellar tendinopathy. *Scand J Med Sci Sports*.23:e89
- 442 30. Vogel HG (1980) Influence of maturation and aging on mechanical and
443 biochemical properties of connective tissue in rats. *Mech Ageing Dev*.14:283
- 444 31. O'Brien TD, Reeves ND, Baltzopoulos V, Jones DA, Maganaris CN (2010)
445 Mechanical properties of the patellar tendon in adults and children. *J*
446 *Biomech*.43:1190
- 447 32. Cassel M, Baur H, Hirschmuller A, Carlsohn A, Frohlich K, Mayer F (2015)
448 Prevalence of Achilles and patellar tendinopathy and their association to
449 intratendinous changes in adolescent athletes. *Scand J Med Sci Sports*.25:e310
- 450 33. Onambele GL, Narici MV, Maganaris CN (2006) Calf muscle-tendon
451 properties and postural balance in old age. *Journal of applied physiology (Bethesda,*
452 *Md : 1985)*.100:2048

- 453 34. Woo SL, Gomez MA, Sites TJ, Newton PO, Orlando CA, Akeson WH (1987)
454 The biomechanical and morphological changes in the medial collateral ligament of
455 the rabbit after immobilization and remobilization. J Bone Joint Surg Am.69:1200
- 456 35. Mo J, Xu H, Qiang B, Giambini H, Kinnick R, An KN, Chen S, Luo Z (2016)
457 Bias of shear wave elasticity measurements in thin layer samples and a simple
458 correction strategy. Springerplus.5:1341
- 459 36. Neviaser A, Andarawis-Puri N, Flatow E (2012) Basic mechanisms of tendon
460 fatigue damage. J Shoulder Elbow Surg.21:158
- 461